



CERTIFICATE OF COMPLIANCE

LRQA

hereby declares that the certification assessment has demonstrated that

COMPACT DRY SL

for the detection of *Salmonella*

Manufactured and supplied by:
Shimadzu Diagnostics Corporation
20th Floor Ueno Frontier Tower
3-24-6 Ueno, Taito-ku, Tokyo, 110-8736
JAPAN

has been validated and revealed to be at least equivalent to the reference method as demonstrated by the validation study report. The summary of the validation report is available on the MicroVal website: www.microval.org

Reference method: ISO 6579-1:2017+A1:2020 Microbiology of the food chain — Horizontal method for the detection, enumeration and serotyping of *Salmonella* Part 1: Detection of *Salmonella* spp.

Scope: Ready to eat and ready to reheat meat and poultry products, eggs and egg products, environmental samples

The validation and certification has been performed in accordance with EN ISO 16140-2:2016 and the MicroVal Rules and Certification Scheme version 9.1.

Certificate no.: 2022LR110

First approval date: 10 November 2023

Expiry date: 9 November 2027

ISSUED BY:

LRQA Nederland B.V.
Rotterdam, The Netherlands